

Surrogate Summary

SDG No.: Q3015

Client: Alliance Technical Group, LLC - Newark

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3015-01	WP0925-PT-VOA-WP	1,2-Dichloroethane-d4	50	51.3	103		74	125
		Dibromofluoromethane	50	49.3	99		75	124
		Toluene-d8	50	49.5	99		86	113
		4-Bromofluorobenzene	50	49.3	99		77	121
VX0916WBL01	VX0916WBL01	1,2-Dichloroethane-d4	50	56.2	112		74	125
		Dibromofluoromethane	50	49.5	99		75	124
		Toluene-d8	50	49.6	99		86	113
		4-Bromofluorobenzene	50	58.3	117		77	121
VX0916WBS01	VX0916WBS01	1,2-Dichloroethane-d4	50	55.7	111		74	125
		Dibromofluoromethane	50	52.0	104		75	124
		Toluene-d8	50	50.9	102		86	113
		4-Bromofluorobenzene	50	53.9	108		77	121
VX0916WBSD0	VX0916WBSD01	1,2-Dichloroethane-d4	50	49.7	99		74	125
		Dibromofluoromethane	50	50.5	101		75	124
		Toluene-d8	50	50.4	101		86	113
		4-Bromofluorobenzene	50	50.0	100		77	121