

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3015	Analytical Method:	SW8260D
Client:	Alliance Technical Group, LLC - New	Datafile :	VX047595.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0916WBS01	Dichlorodifluoromethane	20	18.0	ug/L	90			69	116	
	Chloromethane	20	18.3	ug/L	92			65	116	
	Vinyl chloride	20	18.6	ug/L	93			65	117	
	Ethyl Acetate	20	19.4	ug/L	97			75	115	
	Bromomethane	20	20.6	ug/L	103			58	125	
	Chloroethane	20	19.1	ug/L	96			56	128	
	Trichlorofluoromethane	20	18.8	ug/L	94			73	115	
	1,1,2-Trichlorotrifluoroethane	20	19.3	ug/L	97			80	112	
	1,1-Dichloroethene	20	18.8	ug/L	94			74	110	
	Acrolein	100	94.4	ug/L	94			28	144	
	Acrylonitrile	100	100	ug/L	100			73	113	
	Acetone	100	83.7	ug/L	84			60	125	
	Carbon disulfide	20	18.1	ug/L	91			64	112	
	Methyl tert-butyl Ether	20	20.5	ug/L	103			78	114	
	Methylene Chloride	20	19.1	ug/L	96			72	114	
	trans-1,2-Dichloroethene	20	19.6	ug/L	98			75	108	
	Vinyl Acetate	100	100	ug/L	100			76	115	
	1,1-Dichloroethane	20	19.8	ug/L	99			78	112	
	2-Butanone	100	100	ug/L	100			65	122	
	Carbon Tetrachloride	20	19.1	ug/L	96			77	113	
	2,2-Dichloropropane	20	18.1	ug/L	91			56	164	
	cis-1,2-Dichloroethene	20	19.9	ug/L	100			77	110	
	Chloroform	20	20.3	ug/L	102			79	113	
	1,1,1-Trichloroethane	20	20.2	ug/L	101			80	108	
	1,1-Dichloropropene	20	18.2	ug/L	91			79	110	
	Benzene	20	19.3	ug/L	97			82	109	
	1,2-Dichloroethane	20	19.3	ug/L	97			80	115	
	Trichloroethene	20	19.2	ug/L	96			77	113	
	1,2-Dichloropropane	20	19.5	ug/L	98			83	111	
	Dibromomethane	20	19.3	ug/L	97			82	110	
	Bromodichloromethane	20	19.2	ug/L	96			83	110	
	4-Methyl-2-Pentanone	100	100	ug/L	100			74	118	
	Toluene	20	19.3	ug/L	97			82	110	
	t-1,3-Dichloropropene	20	19.1	ug/L	96			79	110	
	cis-1,3-Dichloropropene	20	19.5	ug/L	98			82	110	
	1,1,2-Trichloroethane	20	19.6	ug/L	98			83	112	
	1,3-Dichloropropane	20	20.2	ug/L	101			83	111	
	2-Chloroethyl vinyl ether	100	84.6	ug/L	85			75	124	
	2-Hexanone	100	100	ug/L	100			73	117	
	Dibromochloromethane	20	19.9	ug/L	100			82	110	
	1,2-Dibromoethane	20	20.3	ug/L	102			81	110	
	Tetrachloroethene	20	18.3	ug/L	92			67	123	
	Chlorobenzene	20	18.5	ug/L	93			82	109	
	1,1,1,2-Tetrachloroethane	20	18.7	ug/L	94			84	111	
	Ethyl Benzene	20	18.7	ug/L	94			83	109	
	m/p-Xylenes	40	37.4	ug/L	94			82	110	
	o-Xylene	20	18.7	ug/L	94			83	109	
	Styrene	20	18.6	ug/L	93			80	111	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3015</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Alliance Technical Group, LLC - New</u>	Datafile :	<u>VX047595.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0916WBS01	Bromoform	20	18.7	ug/L	94			79	109	
	Isopropylbenzene	20	17.9	ug/L	90			83	112	
	1,1,2,2-Tetrachloroethane	20	18.2	ug/L	91			76	118	
	1,2,3-Trichloropropane	20	19.2	ug/L	96			75	112	
	Bromobenzene	20	17.8	ug/L	89			77	116	
	N-propylbenzene	20	17.8	ug/L	89			83	112	
	2-Chlorotoluene	20	18.1	ug/L	91			83	110	
	1,3,5-Trimethylbenzene	20	18.0	ug/L	90			85	112	
	4-Chlorotoluene	20	17.9	ug/L	90			82	110	
	1,2,4-Trimethylbenzene	20	18.0	ug/L	90			85	111	
	Sec-butylbenzene	20	17.9	ug/L	90			81	114	
	p-Isopropyltoluene	20	17.8	ug/L	89			78	116	
	1,3-Dichlorobenzene	20	17.8	ug/L	89			82	108	
	1,4-Dichlorobenzene	20	17.5	ug/L	88			82	107	
	n-Butylbenzene	20	18.0	ug/L	90			75	115	
	1,2-Dichlorobenzene	20	18.1	ug/L	91			82	109	
	1,2-Dibromo-3-Chloropropane	20	18.2	ug/L	91			68	112	
	1,2,4-Trichlorobenzene	20	17.9	ug/L	90			75	113	
	Hexachlorobutadiene	20	17.6	ug/L	88			81	121	
	Naphthalene	20	17.8	ug/L	89			78	119	
	1,2,3-Trichlorobenzene	20	17.8	ug/L	89			76	114	