

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	Q3015	Analytical Method:	SW8260D
Client:	Alliance Technical Group, LLC - New	Datafile :	VX047596.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0916WBSD01	Dichlorodifluoromethane	20	17.8	ug/L	89	1		69	116	19
	Chloromethane	20	17.0	ug/L	85	8		65	116	21
	Vinyl chloride	20	18.3	ug/L	92	1		65	117	19
	Ethyl Acetate	20	19.3	ug/L	97	0		75	115	27
	Bromomethane	20	20.0	ug/L	100	3		58	125	20
	Chloroethane	20	18.6	ug/L	93	3		56	128	20
	Trichlorofluoromethane	20	18.9	ug/L	95	1		73	115	16
	1,1,2-Trichlorotrifluoroethane	20	19.5	ug/L	98	1		80	112	15
	1,1-Dichloroethene	20	19.2	ug/L	96	2		74	110	20
	Acrolein	100	95.1	ug/L	95	1		28	144	30
	Acrylonitrile	100	110	ug/L	110	10		73	113	29
	Acetone	100	82.6	ug/L	83	1		60	125	20
	Carbon disulfide	20	18.0	ug/L	90	1		64	112	20
	Methyl tert-butyl Ether	20	20.0	ug/L	100	3		78	114	20
	Methylene Chloride	20	19.1	ug/L	96	0		72	114	20
	trans-1,2-Dichloroethene	20	19.0	ug/L	95	3		75	108	16
	Vinyl Acetate	100	99.2	ug/L	99	1		76	115	26
	1,1-Dichloroethane	20	19.5	ug/L	98	1		78	112	20
	2-Butanone	100	98.8	ug/L	99	1		65	122	26
	Carbon Tetrachloride	20	19.5	ug/L	98	2		77	113	15
	2,2-Dichloropropane	20	17.6	ug/L	88	3		56	164	20
	cis-1,2-Dichloroethene	20	19.4	ug/L	97	3		77	110	20
	Chloroform	20	20.0	ug/L	100	2		79	113	20
	1,1,1-Trichloroethane	20	20.0	ug/L	100	1		80	108	20
	1,1-Dichloropropene	20	19.5	ug/L	98	7		79	110	16
	Benzene	20	19.5	ug/L	98	1		82	109	15
	1,2-Dichloroethane	20	19.5	ug/L	98	1		80	115	20
	Trichloroethene	20	19.5	ug/L	98	2		77	113	15
	1,2-Dichloropropane	20	19.9	ug/L	100	2		83	111	16
	Dibromomethane	20	20.2	ug/L	101	4		82	110	20
	Bromodichloromethane	20	19.9	ug/L	100	4		83	110	16
	4-Methyl-2-Pentanone	100	110	ug/L	110	10		74	118	25
	Toluene	20	19.8	ug/L	99	2		82	110	16
	t-1,3-Dichloropropene	20	19.8	ug/L	99	3		79	110	20
	cis-1,3-Dichloropropene	20	19.3	ug/L	97	1		82	110	16
	1,1,2-Trichloroethane	20	20.0	ug/L	100	2		83	112	20
	1,3-Dichloropropane	20	20.6	ug/L	103	2		83	111	20
	2-Chloroethyl vinyl ether	100	85.5	ug/L	86	1		75	124	20
	2-Hexanone	100	100	ug/L	100	0		73	117	25
	Dibromochloromethane	20	20.1	ug/L	101	1		82	110	20
	1,2-Dibromoethane	20	20.6	ug/L	103	1		81	110	20
	Tetrachloroethene	20	19.7	ug/L	99	7		67	123	15
	Chlorobenzene	20	19.6	ug/L	98	5		82	109	15
	1,1,1,2-Tetrachloroethane	20	19.7	ug/L	99	5		84	111	20
	Ethyl Benzene	20	19.8	ug/L	99	5		83	109	16
	m/p-Xylenes	40	39.7	ug/L	99	5		82	110	15
	o-Xylene	20	20.1	ug/L	101	7		83	109	20
	Styrene	20	19.9	ug/L	100	7		80	111	17

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3015</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>Alliance Technical Group, LLC - New</u>	Datafile :	<u>VX047596.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0916WBSD01	Bromoform	20	20.2	ug/L	101	7		79	109	20
	Isopropylbenzene	20	20.1	ug/L	101	12		83	112	29
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102	11		76	118	20
	1,2,3-Trichloropropane	20	18.6	ug/L	93	3		75	112	20
	Bromobenzene	20	20.0	ug/L	100	12		77	116	20
	N-propylbenzene	20	20.4	ug/L	102	14		83	112	20
	2-Chlorotoluene	20	20.3	ug/L	102	11		83	110	20
	1,3,5-Trimethylbenzene	20	20.4	ug/L	102	13		85	112	20
	4-Chlorotoluene	20	20.1	ug/L	101	12		82	110	20
	1,2,4-Trimethylbenzene	20	20.4	ug/L	102	13		85	111	20
	Sec-butylbenzene	20	20.8	ug/L	104	14		81	114	20
	p-Isopropyltoluene	20	20.6	ug/L	103	15		78	116	20
	1,3-Dichlorobenzene	20	20.4	ug/L	102	14		82	108	20
	1,4-Dichlorobenzene	20	19.8	ug/L	99	12		82	107	15
	n-Butylbenzene	20	20.9	ug/L	104	14		75	115	20
	1,2-Dichlorobenzene	20	20.3	ug/L	102	11		82	109	20
	1,2-Dibromo-3-Chloropropane	20	20.2	ug/L	101	10		68	112	20
	1,2,4-Trichlorobenzene	20	20.5	ug/L	103	13		75	113	29
	Hexachlorobutadiene	20	21.5	ug/L	108	20		81	121	20
	Naphthalene	20	20.5	ug/L	103	15		78	119	20
	1,2,3-Trichlorobenzene	20	20.9	ug/L	104	16		76	114	29