

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ALLI03
 Lab Code: ACE SDG NO.: Q3015
 Lab File ID: VX047588.D Date Analyzed: 09/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:58
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	173407	5.54	307383	6.75	270040	10.04
UPPER LIMIT	346814	6.044	614766	7.245	540080	10.537
LOWER LIMIT	86703.5	5.044	153692	6.245	135020	9.537
EPA SAMPLE NO.						
WP0925-PT-VOA-WP	166705	5.54	308833	6.75	276674	10.04
VX0916WBL01	170934	5.54	354724	6.75	356396	10.04
VX0916WBS01	184711	5.53	338431	6.75	315471	10.04
VX0916WBSD01	181929	5.54	325791	6.75	290596	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.