

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

|                |                   |                |                   |
|----------------|-------------------|----------------|-------------------|
| Lab Name:      | <u>Alliance</u>   | Contract:      | <u>ALLI03</u>     |
| Lab Code:      | <u>ACE</u>        | SDG NO.:       | <u>Q3015</u>      |
| Lab File ID:   | <u>VX047588.D</u> | Date Analyzed: | <u>09/16/2025</u> |
| Instrument ID: | <u>MSVOA_X</u>    | Time Analyzed: | <u>10:58</u>      |
| GC Column:     | <u>DB-624UI</u>   | ID:            | <u>0.18</u> (mm)  |
|                |                   | Heated Purge:  | (Y/N) <u>N</u>    |

|                  | IS4<br>AREA # | RT #   |  |  |  |  |
|------------------|---------------|--------|--|--|--|--|
| 12 HOUR STD      | 127732        | 12.00€ |  |  |  |  |
| UPPER LIMIT      | 255464        | 12.50€ |  |  |  |  |
| LOWER LIMIT      | 63866         | 11.50€ |  |  |  |  |
| EPA SAMPLE NO.   |               |        |  |  |  |  |
| WP0925-PT-VOA-WP | 133950        | 12.01  |  |  |  |  |
| VX0916WBL01      | 185546        | 12.01  |  |  |  |  |
| VX0916WBS01      | 158261        | 12.01  |  |  |  |  |
| VX0916WBSD01     | 138945        | 12.01  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.