



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN090425\
Data File : VN087713.D
Acq On : 04 Sep 2025 10:10
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N090425W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Fri Sep 05 03:29:53 2025



AutoFind: Averaged scan 1842 to 1858; Bkg corrected with scan 1841

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.1	3426	PASS
75	95	30	60	57.3	8489	PASS
95	95	100	100	100.0	14815	PASS
96	95	5	9	6.8	1008	PASS
173	174	0.00	2	1.3	131	PASS
174	95	50	100	67.0	9919	PASS
175	174	5	9	8.1	802	PASS
176	174	95	101	98.3	9751	PASS
177	176	5	9	7.1	696	PASS