

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN090425

Review By	Mahesh Dadoda	Review On	9/5/2025 2:55:23 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/5/2025 2:57:06 PM
SubDirectory	VN090425	HP Acquire Method	HP Processing Method 624N080525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135371,VP135374
Initial Calibration Stds	VP135375,VP135376,VP135377,VP135378,VP135379
CCC	VP135372,VP135373
Internal Standard/PEM	
ICV/I.BLK	VP135380
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087713.D	04 Sep 2025 10:10		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087714.D	04 Sep 2025 10:54		JC\MD	Ok,M
3	VN0904WBS01	VN0904WBS01	VN087715.D	04 Sep 2025 12:04		JC\MD	Ok,M
4	VN0904WBSD01	VN0904WBSD01	VN087716.D	04 Sep 2025 12:38		JC\MD	Ok,M
5	VN0904WBL01	VN0904WBL01	VN087717.D	04 Sep 2025 12:59		JC\MD	Ok
6	Q3017-01	BP-TB-20250903	VN087718.D	04 Sep 2025 13:35	vial A pH<2 TB	JC\MD	Ok
7	Q3017-02	TT-076-IDWGW-20250	VN087719.D	04 Sep 2025 13:56	vial A+B pH<2	JC\MD	Ok
8	Q3017-04	TT-078-IDWGW-20250	VN087720.D	04 Sep 2025 14:16	vial A+B pH<2	JC\MD	Ok
9	Q3017-05	TT-079-IDWGW-20250	VN087721.D	04 Sep 2025 14:37	vial A+B pH<2	JC\MD	Ok
10	Q3017-03	TT-077-IDWGW-20250	VN087722.D	04 Sep 2025 14:58	vial A+B pH<2	JC\MD	Ok
11	VSTDCCC020	VSTDCCC020EC	VN087723.D	04 Sep 2025 15:21		JC\MD	Ok,M
12	BLK	BLK	VN087724.D	04 Sep 2025 15:41		JC\MD	Ok
13	VSTDICC005	VSTDICC005	VN087725.D	04 Sep 2025 16:23		JC\MD	Ok,M
14	VSTDICCC020	VSTDICCC020	VN087726.D	04 Sep 2025 16:44		JC\MD	Ok,M
15	VSTDICC050	VSTDICC050	VN087727.D	04 Sep 2025 17:05		JC\MD	Ok,M
16	VSTDICC100	VSTDICC100	VN087728.D	04 Sep 2025 17:25		JC\MD	Ok,M
17	VSTDICC150	VSTDICC150	VN087729.D	04 Sep 2025 17:46		JC\MD	Ok,M
18	IBLK	IBLK	VN087730.D	04 Sep 2025 18:07		JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN090425

Review By	Maresh Dadoda	Review On	9/5/2025 2:55:23 PM
Supervise By	Semsettin Yesilyurt	Supervise On	9/5/2025 2:57:06 PM
SubDirectory	VN090425	HP Acquire Method	HP Processing Method 624N080525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP135371,VP135374
Initial Calibration Stds	VP135375,VP135376,VP135377,VP135378,VP135379
CCC	VP135372,VP135373
Internal Standard/PEM	
ICV/I.BLK	VP135380
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	VSTDICV020	ICVVN090425	VN087731.D	04 Sep 2025 18:28		JC\MD	Ok,M
----	------------	-------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN091725

Review By	John Carlone	Review On	9/18/2025 8:19:51 AM
Supervise By	Maresh Dadoda	Supervise On	9/18/2025 3:25:32 PM
SubDirectory	VN091725	HP Acquire Method	HP Processing Method 624N090425W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135514		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135515,VP135516		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087868.D	17 Sep 2025 08:24		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087869.D	17 Sep 2025 09:42	pH#Lot#V12668	JC\MD	Ok,M
3	VN0917WBS01	VN0917WBS01	VN087870.D	17 Sep 2025 10:55		JC\MD	Ok,M
4	VN0917WBSD01	VN0917WBSD01	VN087871.D	17 Sep 2025 11:30		JC\MD	Ok,M
5	VN0917WBL01	VN0917WBL01	VN087872.D	17 Sep 2025 12:03		JC\MD	Ok
6	Q3015-02	WP0925-PT-VOA-WP	VN087873.D	17 Sep 2025 12:38		JC\MD	Ok
7	IBLK	IBLK	VN087874.D	17 Sep 2025 13:11		JC\MD	Ok
8	Q3120-01	OUTFALL-DSN-001	VN087875.D	17 Sep 2025 13:32	vial A pH<2	JC\MD	Ok
9	Q3120-02	OUTFALL-DSN-002	VN087876.D	17 Sep 2025 13:53	vial A pH<2	JC\MD	Ok,M
10	IBLK	IBLK	VN087877.D	17 Sep 2025 14:14		JC\MD	Ok
11	Q3098-01	EFF-WW	VN087878.D	17 Sep 2025 14:35	vial A pH<2 Surrogate fail	JC\MD	ReRun
12	Q3098-02MS	EFF-WWMS	VN087879.D	17 Sep 2025 14:57	vial A pH<2	JC\MD	Ok,M
13	Q3098-03MSD	EFF-WWMSD	VN087880.D	17 Sep 2025 15:18	vial A pH<2	JC\MD	Ok,M
14	IBLK	IBLK	VN087881.D	17 Sep 2025 15:39		JC\MD	Ok
15	Q3098-01	EFF-WW	VN087882.D	17 Sep 2025 16:00	vial B pH<2	JC\MD	Ok,M
16	VSTDCCC020	VSTDCCC020EC	VN087883.D	17 Sep 2025 16:21		JC\MD	Ok,M

M : Manual Integration