

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

METALS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3016

MATRIX: Water

METHOD: 200.8

	NA	NO	YES
1. Calibration Summary met criteria.			✓
2. ICP Interference Check Sample Results Summary Submitted.			✓
3. Laboratory Control Sample Summary (if applicable) Submitted.			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
6. Analysis Holding Time Met			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Holding Times were met for all analysis.			

ADDITIONAL COMMENTS: Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

QA REVIEW

Date