

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3016

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169747BL		WATER		Batch Number:	PB169747		Prep Date:	10/02/2025	
	Mercury	0.027	<0.2	U	0.20	CV	10/02/2025	13:01	LB137392

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3016

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169563BL	WATER			Batch Number:	PB169563		Prep Date:	09/23/2025	
	Aluminum	30.8	<50	U	100	P	09/26/2025	13:49	LB137342
	Antimony	5.44	<25	U	50.0	P	09/26/2025	13:49	LB137342
	Arsenic	3.70	<10	U	20.0	P	09/26/2025	13:49	LB137342
	Barium	19.5	<50	U	100	P	09/26/2025	13:49	LB137342
	Beryllium	0.54	<3	U	6.00	P	09/26/2025	13:49	LB137342
	Boron	13.8	<50	U	100	P	09/26/2025	13:49	LB137342
	Cadmium	0.62	<3	U	6.00	P	09/26/2025	13:49	LB137342
	Calcium	173	<1000	U	2000	P	09/26/2025	13:49	LB137342
	Chromium	1.06	<5	U	10.0	P	09/26/2025	13:49	LB137342
	Cobalt	2.32	<15	U	30.0	P	09/26/2025	13:49	LB137342
	Copper	3.78	<10	U	20.0	P	09/26/2025	13:49	LB137342
	Iron	36.0	<50	U	100	P	09/26/2025	13:49	LB137342
	Lead	2.42	<6	U	12.0	P	09/26/2025	13:49	LB137342
	Magnesium	208	<1000	U	2000	P	09/26/2025	13:49	LB137342
	Manganese	4.92	<10	U	20.0	P	09/26/2025	13:49	LB137342
	Molybdenum	17.6	<100	U	200	P	09/26/2025	13:49	LB137342
	Nickel	3.80	<20	U	40.0	P	09/26/2025	13:49	LB137342
	Potassium	618	<1000	U	2000	P	09/26/2025	13:49	LB137342
	Selenium	5.04	<10	U	20.0	P	09/26/20		