

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3016

Contract: ALLI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169562BS							
Aluminum	ug/L	10000	10300		103	85 - 115	P
Antimony	ug/L	500	499		100	85 - 115	P
Arsenic	ug/L	500	501		100	85 - 115	P
Barium	ug/L	2500	2480		99	85 - 115	P
Beryllium	ug/L	500	495		99	85 - 115	P
Cadmium	ug/L	500	505		101	85 - 115	P
Calcium	ug/L	50000	49000		98	85 - 115	P
Chromium	ug/L	500	501		100	85 - 115	P
Cobalt	ug/L	500	500		100	85 - 115	P
Copper	ug/L	5000	5010		100	85 - 115	P
Iron	ug/L	25000	26300		105	85 - 115	P
Lead	ug/L	2500	2480		99	85 - 115	P
Magnesium	ug/L	50000	49700		99	85 - 115	P
Manganese	ug/L	5000	4930		99	85 - 115	P
Nickel	ug/L	500	497		99	85 - 115	P
Potassium	ug/L	25000	25300		101	85 - 115	P
Tin	ug/L	500	499		100	85 - 115	P
Molybdenum	ug/L	5000	4940		99	85 - 115	P
Selenium	ug/L	500	509		102	85 - 115	P
Silver	ug/L	500	503		101	85 - 115	P
Sodium	ug/L	50000	51300		103	85 - 115	P
Thallium	ug/L	500	504		101	85 - 115	P
Vanadium	ug/L	500	500		100	85 - 115	P
Zinc	ug/L	5000	5130		103	85 - 115	P
Titanium	ug/L	5000	4830		97	85 - 115	P