

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q3016

Instrument ID: P7

Start Date : 09/25/2025

Run Number: LB137409

End Date : 09/25/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Non-Collision Cell |   |                  |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-----------------------------------------------|---|------------------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>165Ho                              | Q | Element<br>209Bi | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1551 | 100                                           |   | 100              |   |         |   |         |   |         |   |
| S2              | S2                 | 1554 | 100                                           |   | 97               |   |         |   |         |   |         |   |
| S3              | S3                 | 1558 | 98                                            |   | 96               |   |         |   |         |   |         |   |
| S4              | S4                 | 1601 | 97                                            |   | 96               |   |         |   |         |   |         |   |
| S5              | S5                 | 1604 | 95                                            |   | 94               |   |         |   |         |   |         |   |
| S6              | S6                 | 1607 | 92                                            |   | 91               |   |         |   |         |   |         |   |
| S7              | S7                 | 1609 | 87                                            |   | 86               |   |         |   |         |   |         |   |
| S8              | S8                 | 1612 | 101                                           |   | 91               |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1620 | 103                                           |   | 101              |   |         |   |         |   |         |   |
| LLICV01         | LLICV01            | 1623 | 101                                           |   | 100              |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1626 | 99                                            |   | 97               |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1629 | 100                                           |   | 96               |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1632 | 103                                           |   | 97               |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1635 | 96                                            |   | 90               |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1638 | 101                                           |   | 99               |   |         |   |         |   |         |   |
| CRI             | CRI                | 1642 | 102                                           |   | 102              |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1645 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1648 |                                               |   |                  |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1651 |                                               |   |                  |   |         |   |         |   |         |   |
| Q3016-02        | WP0925-PT-TM       | 1654 | 104                                           |   | 105              |   |         |   |         |   |         |   |
| Q3016-04        | WP0925-PT-SM       | 1657 | 103                                           |   | 101              |   |         |   |         |   |         |   |
| Q3016-02        | WP0925-PT-TM       | 1700 | 104                                           |   | 105              |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1703 |                                               |   |                  |   |         |   |         |   |         |   |
| Q3016-06        | WP0925-PT-MI       | 1707 | 97                                            |   | 93               |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1710 |                                               |   |                  |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1713 | 95                                            |   | 89               |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1721 | 102                                           |   | 101              |   |         |   |         |   |         |   |
| Q3016-06        | WP0925-PT-MI       | 1724 | 126                                           |   | 121              |   |         |   |         |   |         |   |
| PB169562BL      | PB169562BL         | 1735 | 100                                           |   | 98               |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1738 |                                               |   |                  |   |         |   |         |   |         |   |
| PB169562BS      | PB169562BS         | 1741 | 95                                            |   | 91               |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1744 |                                               |   |                  |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1747 | 94                                            |   | 88               |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1754 | 101                                           |   | 99               |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q3016

Instrument ID: P7

Start Date : 09/25/2025

Run Number: LB137409

End Date : 09/25/2025

| Lab<br>SampleID | Client<br>SampleID | Time | Internal Standard %RI For: Collision Cell |   |         |   |         |   |         |   |         |   |
|-----------------|--------------------|------|-------------------------------------------|---|---------|---|---------|---|---------|---|---------|---|
|                 |                    |      | Element<br>209Bi                          | Q | Element | Q | Element | Q | Element | Q | Element | Q |
| S0              | S0                 | 1551 | 100                                       |   |         |   |         |   |         |   |         |   |
| S2              | S2                 | 1554 | 100                                       |   |         |   |         |   |         |   |         |   |
| S3              | S3                 | 1558 | 98                                        |   |         |   |         |   |         |   |         |   |
| S4              | S4                 | 1601 | 98                                        |   |         |   |         |   |         |   |         |   |
| S5              | S5                 | 1604 | 96                                        |   |         |   |         |   |         |   |         |   |
| S6              | S6                 | 1607 | 92                                        |   |         |   |         |   |         |   |         |   |
| S7              | S7                 | 1609 | 86                                        |   |         |   |         |   |         |   |         |   |
| S8              | S8                 | 1612 | 92                                        |   |         |   |         |   |         |   |         |   |
| ICV01           | ICV01              | 1620 | 101                                       |   |         |   |         |   |         |   |         |   |
| LLICV01         | LLICV01            | 1623 | 100                                       |   |         |   |         |   |         |   |         |   |
| ICB01           | ICB01              | 1626 | 99                                        |   |         |   |         |   |         |   |         |   |
| ICSA01          | ICSA01             | 1629 | 97                                        |   |         |   |         |   |         |   |         |   |
| ICSAB01         | ICSAB01            | 1632 | 98                                        |   |         |   |         |   |         |   |         |   |
| CCV01           | CCV01              | 1635 | 89                                        |   |         |   |         |   |         |   |         |   |
| CCB01           | CCB01              | 1638 | 100                                       |   |         |   |         |   |         |   |         |   |
| CRI             | CRI                | 1642 | 103                                       |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1645 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1648 |                                           |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1651 |                                           |   |         |   |         |   |         |   |         |   |
| Q3016-02        | WP0925-PT-TM       | 1654 | 104                                       |   |         |   |         |   |         |   |         |   |
| Q3016-04        | WP0925-PT-SM       | 1657 | 103                                       |   |         |   |         |   |         |   |         |   |
| Q3016-02        | WP0925-PT-TM       | 1700 | 103                                       |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1703 |                                           |   |         |   |         |   |         |   |         |   |
| Q3016-06        | WP0925-PT-MI       | 1707 | 96                                        |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1710 |                                           |   |         |   |         |   |         |   |         |   |
| CCV02           | CCV02              | 1713 | 91                                        |   |         |   |         |   |         |   |         |   |
| CCB02           | CCB02              | 1721 | 101                                       |   |         |   |         |   |         |   |         |   |
| Q3016-06        | WP0925-PT-MI       | 1724 | 121                                       |   |         |   |         |   |         |   |         |   |
| PB169562BL      | PB169562BL         | 1735 | 102                                       |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1738 |                                           |   |         |   |         |   |         |   |         |   |
| PB169562BS      | PB169562BS         | 1741 | 91                                        |   |         |   |         |   |         |   |         |   |
| ZZZZZZ          | ZZZZZZ             | 1744 |                                           |   |         |   |         |   |         |   |         |   |
| CCV03           | CCV03              | 1747 | 90                                        |   |         |   |         |   |         |   |         |   |
| CCB03           | CCB03              | 1754 | 99                                        |   |         |   |         |   |         |   |         |   |

Internal Standard %RI Limit: 70 -130