



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	109	100	109	65 - 135	P	09/05/2025	11:59	LB137103
	Antimony	52.2	50.0	104	65 - 135	P	09/05/2025	11:59	LB137103
	Arsenic	19.5	20.0	98	65 - 135	P	09/05/2025	11:59	LB137103
	Barium	99.4	100	99	65 - 135	P	09/05/2025	11:59	LB137103
	Beryllium	6.32	6.0	105	65 - 135	P	09/05/2025	11:59	LB137103
	Cadmium	6.07	6.0	101	65 - 135	P	09/05/2025	11:59	LB137103
	Calcium	2070	2000	104	65 - 135	P	09/05/2025	11:59	LB137103
	Chromium	10.3	10.0	103	65 - 135	P	09/05/2025	11:59	LB137103
	Cobalt	30.2	30.0	101	65 - 135	P	09/05/2025	11:59	LB137103
	Copper	21.8	20.0	109	65 - 135	P	09/05/2025	11:59	LB137103
	Iron	99.2	100	99	65 - 135	P	09/05/2025	11:59	LB137103
	Lead	12.3	12.0	102	65 - 135	P	09/05/2025	11:59	LB137103
	Magnesium	2190	2000	110	65 - 135	P	09/05/2025	11:59	LB137103
	Manganese	22.2	20.0	111	65 - 135	P	09/05/2025	11:59	LB137103
	Nickel	40.8	40.0	102	65 - 135	P	09/05/2025	11:59	LB137103
	Potassium	1870	2000	94	65 - 135	P	09/05/2025	11:59	LB137103
	Selenium	19.4	20.0	97	65 - 135	P	09/05/2025	11:59	LB137103
	Silver	10.6	10.0	106	65 - 135	P	09/05/2025	11:59	LB137103
	Sodium	1870	2000	94	65 - 135	P	09/05/2025	11:59	LB137103
	Thallium	41.8	40.0	104	65 - 135	P	09/05/2025	11:59	LB137103
	Vanadium	40.0	40.0	100	65 - 135	P	09/05/2025	11:59	LB137103
	Zinc	42.3	40.0	106	65 - 135	P	09/05/2025	11:59	LB137103
CRA	Mercury	0.23	0.2	114	70 - 130	CV	09/05/2025	15:29	lb137094