



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB21	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	09/05/2025	15:23	lb137094

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB60	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	09/05/2025	15:27	lb137094
CCB61	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	09/05/2025	15:57	lb137094
CCB62	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	09/05/2025	16:27	lb137094
CCB63	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	09/05/2025	16:43	lb137094

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	80.0	100	P	09/05/2025	11:54	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	11:54	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	11:54	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	11:54	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	11:54	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	11:54	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	11:54	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	11:54	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	11:54	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	11:54	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	11:54	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	11:54	LB137103
	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	11:54	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	11:54	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	11:54	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	11:54	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	11:54	LB137103
	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	11:54	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	11:54	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	11:54	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	11:54	LB137103
	Zinc	16.7	+/-20	U	15.0	40.0	P	09/05/2025	11:54	LB137103

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	80.0	100	P	09/05/2025	12:32	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	12:32	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	12:32	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	12:32	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	12:32	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	12:32	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	12:32	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	12:32	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	12:32	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	12:32	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	12:32	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	12:32	LB137103
	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	12:32	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	12:32	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	12:32	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	12:32	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	12:32	LB137103
	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	12:32	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	12:32	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	12:32	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	12:32	LB137103
	Zinc	18.9	+/-20	J	15.0	40.0	P	09/05/2025	12:32	LB137103
CCB02	Aluminum	11.3	+/-50	U	80.0	100	P	09/05/2025	13:24	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	13:24	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	13:24	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	13:24	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	13:24	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	13:24	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	13:24	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	13:24	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	13:24	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	13:24	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	13:24	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	13:24	LB137103
	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	13:24	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	13:24	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	13:24	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	13:24	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	13:24	LB137103

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	13:24	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	13:24	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	13:24	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	13:24	LB137103
	Zinc	16.7	+/-20	U	15.0	40.0	P	09/05/2025	13:24	LB137103
CCB03	Aluminum	11.3	+/-50	U	80.0	100	P	09/05/2025	14:33	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	14:33	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	14:33	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	14:33	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	14:33	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	14:33	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	14:33	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	14:33	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	14:33	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	14:33	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	14:33	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	14:33	LB137103
	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	14:33	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	14:33	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	14:33	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	14:33	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	14:33	LB137103
	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	14:33	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	14:33	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	14:33	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	14:33	LB137103
	Zinc	16.7	+/-20	U	15.0	40.0	P	09/05/2025	14:33	LB137103
CCB04	Aluminum	16.5	+/-50	J	80.0	100	P	09/05/2025	15:40	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	15:40	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	15:40	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	15:40	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	15:40	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	15:40	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	15:40	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	15:40	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	15:40	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	15:40	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	15:40	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	15:40	LB137103

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	15:40	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	15:40	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	15:40	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	15:40	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	15:40	LB137103
	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	15:40	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	15:40	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	15:40	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	15:40	LB137103
	Zinc	16.7	+/-20	U	15.0	40.0	P	09/05/2025	15:40	LB137103
CCB05	Aluminum	11.3	+/-50	U	80.0	100	P	09/05/2025	16:15	LB137103
	Antimony	6.76	+/-25	U	12.5	50.0	P	09/05/2025	16:15	LB137103
	Arsenic	5.12	+/-10	U	15.0	20.0	P	09/05/2025	16:15	LB137103
	Barium	14.6	+/-50	U	25.0	100	P	09/05/2025	16:15	LB137103
	Beryllium	0.56	+/-3	U	1.50	6.00	P	09/05/2025	16:15	LB137103
	Cadmium	0.50	+/-3	U	1.50	6.00	P	09/05/2025	16:15	LB137103
	Calcium	234	+/-1000	U	500	2000	P	09/05/2025	16:15	LB137103
	Chromium	2.12	+/-5	U	5.00	10.0	P	09/05/2025	16:15	LB137103
	Cobalt	2.26	+/-15	U	7.50	30.0	P	09/05/2025	16:15	LB137103
	Copper	4.60	+/-10	U	16.0	20.0	P	09/05/2025	16:15	LB137103
	Iron	23.4	+/-50	U	80.0	100	P	09/05/2025	16:15	LB137103
	Lead	2.30	+/-6	U	9.60	12.0	P	09/05/2025	16:15	LB137103
	Magnesium	244	+/-1000	U	500	2000	P	09/05/2025	16:15	LB137103
	Manganese	5.94	+/-10	U	15.0	20.0	P	09/05/2025	16:15	LB137103
	Nickel	3.06	+/-20	U	10.0	40.0	P	09/05/2025	16:15	LB137103
	Potassium	918	+/-1000	U	1600	2000	P	09/05/2025	16:15	LB137103
	Selenium	9.64	+/-10	U	16.0	20.0	P	09/05/2025	16:15	LB137103
	Silver	1.62	+/-5	U	5.00	10.0	P	09/05/2025	16:15	LB137103
	Sodium	868	+/-1000	U	1000	2000	P	09/05/2025	16:15	LB137103
	Thallium	4.38	+/-20	U	20.0	40.0	P	09/05/2025	16:15	LB137103
	Vanadium	6.26	+/-20	U	20.0	40.0	P	09/05/2025	16:15	LB137103
	Zinc	16.7	+/-20	U	15.0	40.0	P	09/05/2025	16:15	LB137103