

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169582BL		WATER		Batch Number:		PB169582		Prep Date:	09/05/2025	
	Mercury	0.076	<0.2	U	0.16	0.20	CV	09/05/2025	15:39	lb137094

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169555BL	WATER			Batch Number:		PB169555		Prep Date:	09/04/2025	
	Aluminum	5.67	<25	U	40.0	50.0	P	09/05/2025	12:50	LB137103
	Antimony	3.38	<12.5	U	6.25	25.0	P	09/05/2025	12:50	LB137103
	Arsenic	2.56	<5	U	7.50	10.0	P	09/05/2025	12:50	LB137103
	Barium	7.28	<25	U	12.5	50.0	P	09/05/2025	12:50	LB137103
	Beryllium	0.28	<1.5	U	0.75	3.00	P	09/05/2025	12:50	LB137103
	Cadmium	0.25	<1.5	U	0.75	3.00	P	09/05/2025	12:50	LB137103
	Calcium	117	<500	U	250	1000	P	09/05/2025	12:50	LB137103
	Chromium	1.06	<2.5	U	2.50	5.00	P	09/05/2025	12:50	LB137103
	Cobalt	1.13	<7.5	U	3.75	15.0	P	09/05/2025	12:50	LB137103
	Copper	2.30	<5	U	8.00	10.0	P	09/05/2025	12:50	LB137103
	Iron	11.7	<25	U	40.0	50.0	P	09/05/2025	12:50	LB137103
	Lead	1.15	<3	U	4.80	6.00	P	09/05/2025	12:50	LB137103
	Magnesium	122	<500	U	250	1000	P	09/05/2025	12:50	LB137103
	Manganese	2.97	<5	U	7.50	10.0	P	09/05/2025	12:50	LB137103
	Nickel	1.53	<10	U	5.00	20.0	P	09/05/2025	12:50	LB137103
	Potassium	459	<500	U	800	1000	P	09/05/2025	12:50	LB137103
	Selenium	4.82	<5	U	8.00	10.0	P	09/05/2025	12:50	LB137103
	Silver	0.81	<2.5	U	2.50	5.00	P	09/05/2025	12:50	LB137103
	Sodium	434	<500	U	500	1000	P	09/05/2025	12:50	LB137103
	Thallium	2.19	<10	U	10.0	20.0	P	09/05/2025	12:50	LB137103
	Vanadium	3.13	<10	U	10.0	20.0	P	09/05/2025	12:50	LB137103
	Zinc	8.33	<10	U	7.50	20.0	P	09/05/2025	12:50	LB137103