

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169555BS							
Aluminum	ug/L	1000	971		97	86 - 115	P
Antimony	ug/L	400	396		99	88 - 113	P
Arsenic	ug/L	400	380		95	87 - 113	P
Barium	ug/L	100	94.2		94	88 - 113	P
Beryllium	ug/L	100	97.7		98	89 - 112	P
Cadmium	ug/L	100	93.4		93	88 - 113	P
Calcium	ug/L	500	469	J	94	87 - 113	P
Chromium	ug/L	200	194		97	90 - 113	P
Cobalt	ug/L	100	94.4		94	89 - 114	P
Copper	ug/L	150	150		100	86 - 114	P
Iron	ug/L	1500	1450		97	87 - 115	P
Lead	ug/L	500	463		93	86 - 113	P
Magnesium	ug/L	1000	972	J	97	85 - 113	P
Manganese	ug/L	100	98.3		98	90 - 114	P
Nickel	ug/L	250	238		95	88 - 113	P
Potassium	ug/L	5000	4730		95	86 - 114	P
Selenium	ug/L	1000	964		96	83 - 114	P
Silver	ug/L	37.5	36.8		98	84 - 115	P
Sodium	ug/L	1500	1400		93	87 - 115	P
Thallium	ug/L	1000	858		86	85 - 114	P
Vanadium	ug/L	150	146		97	90 - 111	P
Zinc	ug/L	100	97.8		98	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: Q3017

Contract: TETR06

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169582BS Mercury	ug/L	4.0	3.80		95	82 - 119	CV