

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

TT-079-IDWGW-20250903L

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **Lb No.:** lb137103 **Lab Sample ID :** Q3017-05L **SDG No.:** Q3017
Matrix (soil/water): Water **Level (low/med):** LOW
Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ- ence	Q	M
	C		C				
Aluminum	21700	D	22200	D	2		P
Antimony	125	UD	625	UD			P
Arsenic	50.0	UD	250	UD			P
Barium	167	JD	1250	UD	100		P
Beryllium	15.0	UD	75.0	UD			P
Cadmium	15.0	UD	75.0	UD			P
Calcium	9470	D	9210	JD	3		P
Chromium	23.2	JD	125	UD	100		P
Cobalt	12.9	JD	375	UD	100		P
Copper	22.9	JD	250	UD	100		P
Iron	9770	D	9250	D	5		P
Lead	23.8	JD	150	UD	100		P
Magnesium	4110	JD	4150	JD	1		P
Manganese	291	D	291	D	0		P
Mercury	0.20	U	1.00	U			CV
Nickel	17.0	JD	500	UD	100		P
Potassium	5000	UD	25000	UD			P
Selenium	50.0	UD	250	UD			P
Silver	25.0	UD	125	UD			P
Sodium	23800	D	21100	JD	12		P
Thallium	100	UD	500	UD			P
Vanadium	22.9	JD	500	UD	100		P
Zinc	46.1	JD	500	UD	100		P