

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113488.D **Time Analyzed:** 23:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.743	4.648	4.848	580.340	500.000	16.1
Aroclor-1016-2	4.762	4.666	4.866	584.230	500.000	16.8
Aroclor-1016-3	4.819	4.724	4.924	586.660	500.000	17.3
Aroclor-1016-4	4.938	4.843	5.043	539.250	500.000	7.9
Aroclor-1016-5	5.194	5.099	5.299	563.830	500.000	12.8
Aroclor-1260-1	6.230	6.136	6.336	576.580	500.000	15.3
Aroclor-1260-2	6.421	6.326	6.526	585.500	500.000	17.1
Aroclor-1260-3	6.786	6.691	6.891	570.590	500.000	14.1
Aroclor-1260-4	7.045	6.951	7.151	557.290	500.000	11.5
Aroclor-1260-5	7.289	7.194	7.394	549.720	500.000	9.9
Decachlorobiphenyl	8.676	8.582	8.782	51.860	50.000	3.7
Tetrachloro-m-xylene	3.659	3.563	3.763	58.380	50.000	16.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113488.D **Time Analyzed:** 23:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.724	4.629	4.829	535.830	500.000	7.2
Aroclor-1016-2	4.742	4.647	4.847	537.790	500.000	7.6
Aroclor-1016-3	4.916	4.822	5.022	516.270	500.000	3.3
Aroclor-1016-4	4.959	4.864	5.064	502.990	500.000	0.6
Aroclor-1016-5	5.171	5.076	5.276	573.220	500.000	14.6
Aroclor-1260-1	6.198	6.104	6.304	500.660	500.000	0.1
Aroclor-1260-2	6.387	6.292	6.492	488.800	500.000	-2.2
Aroclor-1260-3	6.538	6.443	6.643	456.710	500.000	-8.7
Aroclor-1260-4	7.007	6.912	7.112	470.540	500.000	-5.9
Aroclor-1260-5	7.249	7.155	7.355	459.290	500.000	-8.1
Decachlorobiphenyl	8.620	8.528	8.728	55.050	50.000	10.1
Tetrachloro-m-xylene	3.651	3.555	3.755	54.760	50.000	9.5