

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074886.D **Time Analyzed:** 08:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.706	5.906	509.440	500.000	1.9
Aroclor-1016-2	5.827	5.727	5.927	530.120	500.000	6.0
Aroclor-1016-3	5.890	5.790	5.990	478.420	500.000	-4.3
Aroclor-1016-4	5.988	5.888	6.088	508.630	500.000	1.7
Aroclor-1016-5	6.280	6.180	6.380	528.550	500.000	5.7
Aroclor-1260-1	7.397	7.297	7.497	490.460	500.000	-1.9
Aroclor-1260-2	7.650	7.550	7.750	485.610	500.000	-2.9
Aroclor-1260-3	8.008	7.908	8.108	516.920	500.000	3.4
Aroclor-1260-4	8.236	8.136	8.336	496.700	500.000	-0.7
Aroclor-1260-5	8.562	8.462	8.662	476.630	500.000	-4.7
Decachlorobiphenyl	10.433	10.333	10.533	56.080	50.000	12.2
Tetrachloro-m-xylene	4.654	4.554	4.754	53.210	50.000	6.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074886.D **Time Analyzed:** 08:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.799	4.999	555.770	500.000	11.2
Aroclor-1016-2	4.952	4.855	5.055	568.340	500.000	13.7
Aroclor-1016-3	5.072	4.976	5.176	539.900	500.000	8.0
Aroclor-1016-4	5.112	5.017	5.217	528.240	500.000	5.6
Aroclor-1016-5	5.327	5.231	5.431	475.800	500.000	-4.8
Aroclor-1260-1	6.357	6.261	6.461	508.560	500.000	1.7
Aroclor-1260-2	6.542	6.448	6.648	499.680	500.000	-0.1
Aroclor-1260-3	6.696	6.602	6.802	450.490	500.000	-9.9
Aroclor-1260-4	7.166	7.072	7.272	489.810	500.000	-2.0
Aroclor-1260-5	7.405	7.311	7.511	473.770	500.000	-5.2
Decachlorobiphenyl	8.806	8.715	8.915	55.940	50.000	11.9
Tetrachloro-m-xylene	3.794	3.699	3.899	59.900	50.000	19.8