

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074897.D **Time Analyzed:** 12:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.812	5.706	5.906	518.950	500.000	3.8
Aroclor-1016-2	5.833	5.727	5.927	533.100	500.000	6.6
Aroclor-1016-3	5.896	5.790	5.990	487.500	500.000	-2.5
Aroclor-1016-4	5.993	5.888	6.088	512.430	500.000	2.5
Aroclor-1016-5	6.286	6.180	6.380	533.180	500.000	6.6
Aroclor-1260-1	7.404	7.297	7.497	493.310	500.000	-1.3
Aroclor-1260-2	7.656	7.550	7.750	489.630	500.000	-2.1
Aroclor-1260-3	8.014	7.908	8.108	521.850	500.000	4.4
Aroclor-1260-4	8.243	8.136	8.336	497.170	500.000	-0.6
Aroclor-1260-5	8.569	8.462	8.662	482.090	500.000	-3.6
Decachlorobiphenyl	10.442	10.333	10.533	56.640	50.000	13.3
Tetrachloro-m-xylene	4.660	4.554	4.754	53.390	50.000	6.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TETR06
Lab Code: ACE **SDG NO.:** Q3017
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/05/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074897.D **Time Analyzed:** 12:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.799	4.999	542.700	500.000	8.5
Aroclor-1016-2	4.957	4.855	5.055	551.020	500.000	10.2
Aroclor-1016-3	5.077	4.976	5.176	549.490	500.000	9.9
Aroclor-1016-4	5.118	5.017	5.217	522.750	500.000	4.6
Aroclor-1016-5	5.333	5.231	5.431	485.330	500.000	-2.9
Aroclor-1260-1	6.362	6.261	6.461	514.650	500.000	2.9
Aroclor-1260-2	6.548	6.448	6.648	494.340	500.000	-1.1
Aroclor-1260-3	6.702	6.602	6.802	455.980	500.000	-8.8
Aroclor-1260-4	7.172	7.072	7.272	485.530	500.000	-2.9
Aroclor-1260-5	7.411	7.311	7.511	471.710	500.000	-5.7
Decachlorobiphenyl	8.813	8.715	8.915	55.610	50.000	11.2
Tetrachloro-m-xylene	3.800	3.699	3.899	59.960	50.000	19.9