

Surrogate Summary

SDG No.: Q3017

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113002.D	PIBLK-PO113002.D	Tetrachloro-m-xyl	1	20	16.6	83		60	140
		Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	16.9	84		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113482.D	PIBLK-PO113482.D	Tetrachloro-m-xyl	1	20	18.8	94		60	140
		Decachlorobiphen	1	20	17.8	89		60	140
		Tetrachloro-m-xyl	2	20	18.2	91		60	140
		Decachlorobiphen	2	20	19.4	97		60	140
PB169547BL	PB169547BL	Tetrachloro-m-xyl	1	20	20.4	102		35	137
		Decachlorobiphen	1	20	18.9	95		40	135
		Tetrachloro-m-xyl	2	20	19.5	97		35	137
		Decachlorobiphen	2	20	20.6	103		40	135
PB169547BS	PB169547BS	Tetrachloro-m-xyl	1	20	20.2	101		35	137
		Decachlorobiphen	1	20	19.4	97		40	135
		Tetrachloro-m-xyl	2	20	19.2	96		35	137
		Decachlorobiphen	2	20	21.3	106		40	135
PB169547BSD	PB169547BSD	Tetrachloro-m-xyl	1	20	19.6	98		35	137
		Decachlorobiphen	1	20	18.8	94		40	135
		Tetrachloro-m-xyl	2	20	18.6	93		35	137
		Decachlorobiphen	2	20	20.5	102		40	135
I.BLK-PO113492.D	PIBLK-PO113492.D	Tetrachloro-m-xyl	1	20	19.6	98		60	140
		Decachlorobiphen	1	20	18.9	94		60	140
		Tetrachloro-m-xyl	2	20	18.6	93		60	140
		Decachlorobiphen	2	20	20.2	101		60	140
I.BLK-PP074490.D	PIBLK-PP074490.D	Tetrachloro-m-xyl	1	20	17.8	89		60	140
		Decachlorobiphen	1	20	18.2	91		60	140
		Tetrachloro-m-xyl	2	20	15.7	78		60	140
		Decachlorobiphen	2	20	16.9	85		60	140
I.BLK-PP074887.D	PIBLK-PP074887.D	Tetrachloro-m-xyl	1	20	18.0	90		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	16.6	83		60	140
		Decachlorobiphen	2	20	14.3	72		60	140
Q3017-02	TT-076-IDWGW-20250903	Tetrachloro-m-xyl	1	20	20.8	104		35	137
		Decachlorobiphen	1	20	19.3	97		40	135
		Tetrachloro-m-xyl	2	20	20.8	104		35	137
		Decachlorobiphen	2	20	15.9	80		40	135
Q3017-03	TT-077-IDWGW-20250903	Tetrachloro-m-xyl	1	20	18.6	93		35	137
		Decachlorobiphen	1	20	20.2	101		40	135
		Tetrachloro-m-xyl	2	20	18.3	91		35	137
		Decachlorobiphen	2	20	16.9	84		40	135
Q3017-04	TT-078-IDWGW-20250903	Tetrachloro-m-xyl	1	20	20.3	102		35	137

Surrogate Summary

SDG No.: Q3017

Client: Tetra Tech NUS, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q3017-04	TT-078-IDWGW-20250903	Decachlorobiphen	1	20	18.2	91		40	135
		Tetrachloro-m-xyl	2	20	21.4	107		35	137
		Decachlorobiphen	2	20	14.7	73		40	135
Q3017-05	TT-079-IDWGW-20250903	Tetrachloro-m-xyl	1	20	21.0	105		35	137
		Decachlorobiphen	1	20	20.5	103		40	135
		Tetrachloro-m-xyl	2	20	24.9	124		35	137
		Decachlorobiphen	2	20	17.2	86		40	135
I.BLK-PP074898.D	PIBLK-PP074898.D	Tetrachloro-m-xyl	1	20	17.7	88		60	140
		Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	15.6	78		60	140
		Decachlorobiphen	2	20	13.9	70		60	140