



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3017</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>PO113485.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169547BSD (Column 1)	AR1016	5	4.70	ug/L	94	0			46	129	20
	AR1260	5	4.30	ug/L	86	5			45	134	20
PB169547BSD (Column 2)	AR1016	5	4.30	ug/L	86	2			46	129	20
	AR1260	5	3.80	ug/L	76	3			45	134	20