

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/03/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/04/25	
Client Sample ID:	TT-076-IDWGW-20250903		SDG No.:	Q3017	
Lab Sample ID:	Q3017-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087719.D	1	09/04/25 13:56	VN090425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.77	U	0.77	5.00	ug/L
74-87-3	Chloromethane	0.64	U	0.64	5.00	ug/L
75-01-4	Vinyl Chloride	0.83	U	0.83	5.00	ug/L
74-83-9	Bromomethane	0.80	U	0.80	5.00	ug/L
75-00-3	Chloroethane	2.30	U	2.30	5.00	ug/L
75-69-4	Trichlorofluoromethane	0.80	U	0.80	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.96	U	0.96	5.00	ug/L
75-35-4	1,1-Dichloroethene	0.76	U	0.76	5.00	ug/L
67-64-1	Acetone	4.60	U	4.60	25.0	ug/L
75-15-0	Carbon Disulfide	0.82	U	0.82	5.00	ug/L
1634-04-4	Methyl tert-Butyl Ether	0.77	U	0.77	5.00	ug/L
79-20-9	Methyl Acetate	0.91	U	0.91	5.00	ug/L
75-09-2	Methylene Chloride	0.86	U	0.86	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.82	U	0.82	5.00	ug/L
75-34-3	1,1-Dichloroethane	0.68	U	0.68	5.00	ug/L
110-82-7	Cyclohexane	0.92	U	0.92	5.00	ug/L
78-93-3	2-Butanone	2.00	U	2.00	25.0	ug/L
56-23-5	Carbon Tetrachloride	0.74	U	0.74	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.80	U	0.80	5.00	ug/L
67-66-3	Chloroform	0.55	U	0.55	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.63	U	0.63	5.00	ug/L
108-87-2	Methylcyclohexane	0.73	U	0.73	5.00	ug/L
71-43-2	Benzene	0.45	U	0.45	5.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.50	5.00	ug/L
79-01-6	Trichloroethene	6.70		0.49	5.00	ug/L
78-87-5	1,2-Dichloropropane	0.46	U	0.46	5.00	ug/L
75-27-4	Bromodichloromethane	0.64	U	0.64	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	3.00	U	3.00	25.0	ug/L
108-88-3	Toluene	0.46	U	0.46	5.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.72	U	0.72	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/03/25	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/04/25	
Client Sample ID:	TT-076-IDWGW-20250903		SDG No.:	Q3017	
Lab Sample ID:	Q3017-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087719.D	1	09/04/25 13:56	VN090425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-01-5	cis-1,3-Dichloropropene	0.67	U	0.67	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.45	U	0.45	5.00	ug/L
591-78-6	2-Hexanone	3.20	U	3.20	25.0	ug/L
124-48-1	Dibromochloromethane	0.66	U	0.66	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.56	U	0.56	5.00	ug/L
127-18-4	Tetrachloroethene	0.84	U	0.84	5.00	ug/L
108-90-7	Chlorobenzene	0.47	U	0.47	5.00	ug/L
100-41-4	Ethyl Benzene	0.56	U	0.56	5.00	ug/L
179601-23-1	m/p-Xylenes	1.30	U	1.30	10.0	ug/L
95-47-6	o-Xylene	0.67	U	0.67	5.00	ug/L
100-42-5	Styrene	0.72	U	0.72	5.00	ug/L
75-25-2	Bromoform	0.94	U	0.94	5.00	ug/L
98-82-8	Isopropylbenzene	0.78	U	0.78	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.44	U	0.44	5.00	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.59	U	0.59	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.67	U	0.67	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.81	U	0.81	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.67	U	0.67	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.86	U	0.86	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.00	U	1.00	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.10	U	1.10	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	29.8		91 - 110	99%	SPK: 30
2037-26-5	Toluene-d8	29.5		91 - 112	98%	SPK: 30
460-00-4	4-Bromofluorobenzene	25.4		63 - 112	85%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	31600	7.794			
540-36-3	1,4-Difluorobenzene	160000	9.082			
3114-55-4	Chlorobenzene-d5	142000	11.847			

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	09/03/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	09/04/25
Client Sample ID:	TT-076-IDWGW-20250903		SDG No.:	Q3017
Lab Sample ID:	Q3017-02		Matrix:	Water
Analytical Method:	E624.1		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VN087719.D	1	09/04/25 13:56	VN090425

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products