

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
Lab Code: ACE SDG NO.: Q3017
Lab File ID: VN087714.D Date Analyzed: 09/04/2025
Instrument ID: MSVOA_N Time Analyzed: 10:54
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	48350	7.80	251837	9.08	230082	11.85
UPPER LIMIT	96700	8.3	503674	9.583	460164	12.347
LOWER LIMIT	24175	7.3	125919	8.583	115041	11.347
EPA SAMPLE NO.						
BP-TB-20250903	24981	7.80	130931	9.08	117412	11.85
TT-076-IDWGW-20250903	31627	7.79	160432	9.08	142263	11.85
TT-077-IDWGW-20250903	41995	7.79	225576	9.08	201361	11.85
TT-078-IDWGW-20250903	36172	7.80	190299	9.08	168068	11.85
TT-079-IDWGW-20250903	39130	7.79	202717	9.08	174966	11.84
VN0904WBL01	37882	7.79	207176	9.08	183441	11.85
VN0904WBS01	43425	7.79	231247	9.08	210355	11.85
VN0904WBSD01	46987	7.79	246294	9.08	222512	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.