

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG NO.: Q3017
 Lab File ID: VN087714.D Date Analyzed: 09/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:54
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
BP-TB-20250903	0	0.00				
TT-076-IDWGW-20250903	0	0.00				
TT-077-IDWGW-20250903	0	0.00				
TT-078-IDWGW-20250903	0	0.00				
TT-079-IDWGW-20250903	0	0.00				
VN0904WBL01	0	0.00				
VN0904WBS01	0	0.00				
VN0904WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.