

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3018

Contract: ALLI03

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	260000	255000	102	216000	294000	09/26/2025	13:04	LB137342
	Antimony	-2.91			-50	50	09/26/2025	13:04	LB137342
	Arsenic	3.14			-20	20	09/26/2025	13:04	LB137342
	Barium	3.49	6.0	58	-94	106	09/26/2025	13:04	LB137342
	Beryllium	1.67			-6	6	09/26/2025	13:04	LB137342
	Boron	72.7	1000	7	-100	100	09/26/2025	13:04	LB137342
	Cadmium	-1.23	1.0	123	-5	7	09/26/2025	13:04	LB137342
	Calcium	241000	245000	98	208000	282000	09/26/2025	13:04	LB137342
	Chromium	45.7	52.0	88	42	62	09/26/2025	13:04	LB137342
	Cobalt	1.48			-30	30	09/26/2025	13:04	LB137342
	Copper	9.02	2.0	451	-18	22	09/26/2025	13:04	LB137342
	Iron	102000	101000	101	85600	116500	09/26/2025	13:04	LB137342
	Lead	-0.70			-12	12	09/26/2025	13:04	LB137342
	Magnesium	258000	255000	101	216000	294000	09/26/2025	13:04	LB137342
	Manganese	6.38	7.0	91	-13	27	09/26/2025	13:04	LB137342
	Molybdenum	0.17	1000		-200	200	09/26/2025	13:04	LB137342
	Nickel	3.29	2.0	164	-38	42	09/26/2025	13:04	LB137342
	Potassium	43.8			0	0	09/26/2025	13:04	LB137342
	Selenium	7.79			-20	20	09/26/2025	13:04	LB137342
	Silver	-1.79			-10	10	09/26/2025	13:04	LB137342
Sodium	76.3			0	0	09/26/2025	13:04	LB137342	
Strontium	3.74			0	0	09/26/2025	13:04	LB137342	
Thallium	5.89			-40	40	09/26/2025	13:04	LB137342	
Tin	-4.75	1000		-40	40	09/26/2025	13:04	LB137342	
Titanium	-4.04			0	0	09/26/2025	13:04	LB137342	
Vanadium	0.41			-40	40	09/26/2025	13:04	LB137342	
Zinc	2.43			-40	40	09/26/2025	13:04	LB137342	
ICSAB01	Aluminum	249000	247000	101	209000	285000	09/26/2025	13:08	LB137342
	Antimony	609	618	98	525	711	09/26/2025	13:08	LB137342
	Arsenic	108	104	104	88.4	120	09/26/2025	13:08	LB137342
	Barium	495	537	92	437	637	09/26/2025	13:08	LB137342
	Beryllium	504	495	102	420	570	09/26/2025	13:08	LB137342
	Boron	981	1000	98	850	1150	09/26/2025	13:08	LB137342
	Cadmium	983	972	101	826	1120	09/26/2025	13:08	LB137342
	Calcium	236000	235000	100	199000	271000	09/26/2025	13:08	LB137342
	Chromium	544	542	100	460	624	09/26/2025	13:08	LB137342
	Cobalt	494	476	104	404	548	09/26/2025	13:08	LB137342
	Copper	500	511	98	434	588	09/26/2025	13:08	LB137342
	Iron	98300	99300	99	84400	114500	09/26/2025	13:08	LB137342
	Lead	43.2	49.0	88	37	61	09/26/2025	13:08	LB137342
	Magnesium	250000	248000	101	210000	286000	09/26/2025	13:08	LB137342
Manganese	493	507	97	430	584	09/26/2025	13:08	LB137342	

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Alliance Technical Group, LLC - Newark
Contract: ALLI03
ICS Source: EPA

SDG No.: Q3018
Lab Code: ACE
Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Molybdenum	960	1000	96	850	1150	09/26/2025	13:08	LB137342
	Nickel	981	954	103	810	1100	09/26/2025	13:08	LB137342
	Potassium	-20.4			0	0	09/26/2025	13:08	LB137342
	Selenium	60.5	46.0	132	26	66	09/26/2025	13:08	LB137342
	Silver	209	201	104	170	232	09/26/2025	13:08	LB137342
	Sodium	69.1			0	0	09/26/2025	13:08	LB137342
	Strontium	1020			0	0	09/26/2025	13:08	LB137342
	Thallium	113	108	105	68	148	09/26/2025	13:08	LB137342
	Tin	935	1000	94	850	1150	09/26/2025	13:08	LB137342
	Titanium	935			0	0	09/26/2025	13:08	LB137342
	Vanadium	483	491	98	417	565	09/26/2025	13:08	LB137342
	Zinc	1040	952	109	809	1095	09/26/2025	13:08	LB137342