

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q3018

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169560BL	WATER			Batch Number:	PB169560		Prep Date:	09/23/2025	
	Aluminum	1.94	<10	U	20.0	P	09/25/2025	17:38	LB137409
	Antimony	0.11	<1	U	2.00	P	09/25/2025	17:38	LB137409
	Arsenic	0.089	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Barium	0.21	<5	U	10.0	P	09/25/2025	17:38	LB137409
	Beryllium	0.32	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Cadmium	0.34	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Calcium	45.7	<250	U	500	P	09/25/2025	17:38	LB137409
	Chromium	0.21	<1	U	2.00	P	09/25/2025	17:38	LB137409
	Cobalt	0.070	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Copper	0.30	<1	U	2.00	P	09/25/2025	17:38	LB137409
	Iron	7.81	<25	U	50.0	P	09/25/2025	17:38	LB137409
	Lead	0.21	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Magnesium	19.5	<250	U	500	P	09/25/2025	17:38	LB137409
	Manganese	0.43	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Nickel	0.27	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Potassium	36.4	<250	U	500	P	09/25/2025	17:38	LB137409
	Tin	0.31	<2.5	U	5.00	P	09/25/2025	17:38	LB137409
	Molybdenum	0.27	<2.5	U	5.00	P	09/25/2025	17:38	LB137409
	Selenium	2.90	<2.5	U	5.00	P	09/25/2025	17:38	LB137409
	Silver	0.060	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Sodium	128	<250	U	500	P	09/25/2025	17:38	LB137409
	Thallium	0.060	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Vanadium	0.077	<2.5	U	5.00	P	09/25/2025	17:38	LB137409
	Zinc	1.25	<2.5	U	5.00	P	09/25/2025	17:38	LB137409
	Strontium	0.084	<0.5	U	1.00	P	09/25/2025	17:38	LB137409
	Titanium	0.20	<2.5	U	5.00	P	09/25/2025	17:38	LB137409