

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q3018

Instrument ID: P7

Start Date : 09/25/2025

Run Number: LB137409

End Date : 09/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1551	100		100							
S2	S2	1554	100		97							
S3	S3	1558	98		96							
S4	S4	1601	97		96							
S5	S5	1604	95		94							
S6	S6	1607	92		91							
S7	S7	1609	87		86							
S8	S8	1612	101		91							
ICV01	ICV01	1620	103		101							
LLICV01	LLICV01	1623	101		100							
ICB01	ICB01	1626	99		97							
ICSA01	ICSA01	1629	100		96							
ICSAB01	ICSAB01	1632	103		97							
CCV01	CCV01	1635	96		90							
CCB01	CCB01	1638	101		99							
CRI	CRI	1642	102		102							
Q3018-02	WP0925-PT-TM	1645	104		104							
Q3018-04	WP0925-PT-SM	1648	104		105							
Q3018-02	WP0925-PT-TM	1651	105		103							
ZZZZZZ	ZZZZZZ	1654										
ZZZZZZ	ZZZZZZ	1657										
ZZZZZZ	ZZZZZZ	1700										
Q3018-06	WP0925-PT-MI	1703	96		95							
ZZZZZZ	ZZZZZZ	1707										
Q3018-06	WP0925-PT-MI	1710	112		110							
CCV02	CCV02	1713	95		89							
CCB02	CCB02	1721	102		101							
ZZZZZZ	ZZZZZZ	1724										
ZZZZZZ	ZZZZZZ	1735										
PB169560BL	PB169560BL	1738	102		100							
ZZZZZZ	ZZZZZZ	1741										
PB169560BS	PB169560BS	1744	95		93							
CCV03	CCV03	1747	94		88							
CCB03	CCB03	1754	101		99							

Internal Standard %RI Limit: 70 -130

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q3018

Instrument ID: P7

Start Date : 09/25/2025

Run Number: LB137409

End Date : 09/25/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1551	100									
S2	S2	1554	100									
S3	S3	1558	98									
S4	S4	1601	98									
S5	S5	1604	96									
S6	S6	1607	92									
S7	S7	1609	86									
S8	S8	1612	92									
ICV01	ICV01	1620	101									
LLICV01	LLICV01	1623	100									
ICB01	ICB01	1626	99									
ICSA01	ICSA01	1629	97									
ICSAB01	ICSAB01	1632	98									
CCV01	CCV01	1635	89									
CCB01	CCB01	1638	100									
CRI	CRI	1642	103									
Q3018-02	WP0925-PT-TM	1645	102									
Q3018-04	WP0925-PT-SM	1648	102									
Q3018-02	WP0925-PT-TM	1651	102									
ZZZZZZ	ZZZZZZ	1654										
ZZZZZZ	ZZZZZZ	1657										
ZZZZZZ	ZZZZZZ	1700										
Q3018-06	WP0925-PT-MI	1703	96									
ZZZZZZ	ZZZZZZ	1707										
Q3018-06	WP0925-PT-MI	1710	109									
CCV02	CCV02	1713	91									
CCB02	CCB02	1721	101									
ZZZZZZ	ZZZZZZ	1724										
ZZZZZZ	ZZZZZZ	1735										
PB169560BL	PB169560BL	1738	102									
ZZZZZZ	ZZZZZZ	1741										
PB169560BS	PB169560BS	1744	94									
CCV03	CCV03	1747	90									
CCB03	CCB03	1754	99									

Internal Standard %RI Limit: 70 -130