

Analytical Summary Report

Analysis Method: SM2120 B

Parameter: Color

Run Number: LB137195

Reviewed By: Iwona

Supervisor Review By: JIGNESH

pH Meter ID: PH METER-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP114756
Calibration std, 5 color unit	WP114757
Calibration std, 10 color unit	WP114758
Calibration std, 15 color unit	WP114759
Calibration std, 20 color unit	WP114760
Calibration std, 25 color unit	WP114761
Calibration std, 30 color unit	WP114762
Calibration std, 40 color unit	WP114763
Calibration std, 50 color unit	WP114764
Calibration std, 60 color unit	WP114765
Calibration std, 70 color unit	WP114766
Calibration std, 100 color unit	WP114767

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB137195BL	LB137195BL	1	5.02	N/A	0.00	09/15/2025	13:40
2	Q3019-27	WP0925-PT-COL-WP	1	1.12	N/A	50.00	09/15/2025	13:45
3	Q3019-27DUP	WP0925-PT-COL-WPDUP	1	1.11	N/A	50.00	09/15/2025	13:50

WORKLIST(Hardcopy Internal Chain)

LB137195

WorkList Name : COLOR-091525

WorkList ID : 191879

Department : Wet-Chemistry

Date : 09-15-2025 13:10:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3019-27	WP0925-PT-COL-WP	Water	Color	Cool 4 deg C	ALLI03	QA Of	09/02/2025	SM2120 B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

09/15/25 13:20
12 (JC)
SJ (QA)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

NA

it

6