

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB137482

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP115134
Turbidity Calibration std, 0NTU	WP115127
Turbidity Calibration std, 20NTU	WP115129
Turbidity Calibration std, 0.5NTU	WP115133
Turbidity Calibration std, 40NTU	WP115128
Turbidity Calibration std, 1NTU	WP115132
Turbidity Calibration - CCV std, 10 NTU	WP115130
10 NTU Standard 500 ml	W3116

Intercept: 0.0865

Slope: 1.1087

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.06		10/09/2025	15:10
2	CAL2	1	1	1.260	1.06	5.8	10/09/2025	15:13
3	CAL3	5	1	5.670	5.04	0.7	10/09/2025	15:16
4	CAL4	10	1	11.200	10.02	0.2	10/09/2025	15:20
5	CAL5	20	1	21.900	19.68	-1.6	10/09/2025	15:23
6	CAL6	40	1	44.600	40.15	0.4	10/09/2025	15:26

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB137482

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	11.600	10.385	10/09/2025	15:29
2	ICB		1	0.152	0.059	10/09/2025	15:32
3	CCV1	10	1	10.800	9.663	10/09/2025	15:35
4	CCB1		1	0.164	0.070	10/09/2025	15:38
5	LB137482BL		1	0.184	0.088	10/09/2025	15:41
6	Q2893-09		1	0.741	0.590	10/09/2025	15:44
7	Q3019-34		1	14.900	13.361	10/09/2025	15:47
8	CCV2	10	1	10.200	9.122	10/09/2025	15:51
9	CCB2		1	0.174	0.079	10/09/2025	15:54

LB 137482
LB 137483

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY-100925

WorkList ID : 192376

Department : Wet-Chemistry

Date : 10-09-2025 13:57:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3019-33	WP0925-PT-TURB-WP	Water	Turbidity	Cool 4 deg C	ALLI03	QA Of	09/02/2025	SM2130 B
Q3019-34	WP0925-PT-TURB-WP	Water	Turbidity	Cool 4 deg C	ALLI03	QA Of	09/02/2025	180.1

Date/Time 10/09/25 16:10
Raw Sample Received by: 12(WC)
Raw Sample Relinquished by: SJ(QAO)

Date/Time MA
Raw Sample Received by: 11
Raw Sample Relinquished by: (1)