



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**VERINA CONSULTING GROUP, LLC**

**Project Name: Rotor Clip NJ WTD - 2025**

**Project # N/A**

**Order ID # Q3020**

**Test Name: Ammonia,BOD5,COD,Oil and Grease,Residual Chlorine,TSS**

### **A. Number of Samples and Date of Receipt:**

4 Water samples were received on 09/04/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Ammonia,BOD5,COD,Oil and Grease,Residual Chlorine,TSS. This data package contains results for Ammonia,BOD5,COD,Oil and Grease,Residual Chlorine,TSS.

### **C. Analytical Techniques:**

The analysis of Oil and Grease was based on method 1664A, The analysis of TSS was based on method SM2540 D, The analysis of Residual Chlorine was based on method SM4500 Cl G, The analysis of Ammonia was based on method SM4500-NH3, The analysis of BOD5 was based on method SM5210 B and The analysis of COD was based on method SM5220 D.

### **D. QA/ QC Samples:**

The Holding Times were met for all samples except for WATER-TREATMENT DISCHARGE of Residual Chlorine as sample was receive out of holding time.

The Blank Spike met requirements for all compounds.

The Duplicate (WATER-TREATMEN DISCHARGEDUP) analysis met criteria for all compounds except for Residual Chlorine due to sample matrix interference.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

### **E. Additional Comments:**

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_