



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137088

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.419	0.4	105	90-110	09/05/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.379	0.4	95	90-110	09/05/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.419	0.4	105	90-110	09/05/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Project: Rotor Clip NJ WTD - 2025

RunNo.: LB137105

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	09/08/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025

Initial and Continuing Calibration Verification**Client:** VERINA CONSULTING GROUP, LLC**SDG No.:** Q3020**Project:** Rotor Clip NJ WTD - 2025**RunNo.:** LB137109

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: COD	ICV	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: COD	CCV1	mg/L	48.150	50	96	95-105	09/08/2025
Sample ID: COD	CCV2	mg/L	51.184	50	102	95-105	09/08/2025
Sample ID: COD	CCV3	mg/L	50.173	50	100	95-105	09/08/2025