

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3007-02
Client ID:	CompDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
BOD5	mg/L	+/-20	846		827		1	2.23		09/05/2025
TSS	mg/L	+/-5	228		226		1	0.88		09/08/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3020-01
Client ID:	WATER-TREATMEN DISCHARGEDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.035	J	0.045	J	1	25	*	09/05/2025
Ammonia as N	mg/L	+/-20	0.30		0.30		1	0		09/08/2025
COD	mg/L	+/-20	61.3		60.3		1	1.64		09/08/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3020-01
Client ID:	WATER-TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Residual Chlorine	mg/L	+/-20	0.41		0.40		1	2.48		09/05/2025
Ammonia as N	mg/L	+/-20	1.20		1.20		1	0		09/08/2025
COD	mg/L	+/-20	106		107		1	0.94		09/08/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3020-02
Client ID:	WATER-TREATMENT DISCHARGEMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	30.8		30.9		1	0.32		09/05/2025

Duplicate Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Sample ID:	Q3028-01
Client ID:	Q3028-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	30.5		30.6		1	0.33		09/05/2025