

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137083

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137083BS							
Oil and Grease	mg/L	20.0	17.0		85	1	78-114	09/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137088

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137088BS							
Residual Chlorine	mg/L	0.4	0.41		102	1	90-110	09/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137092

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137092BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	09/05/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137104BS							
TSS	mg/L	550	531		96	1	90-110	09/08/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137109

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137109BS							
COD	mg/L	50	51.2		102	1	90-110	09/08/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q3020
Project:	Rotor Clip NJ WTD - 2025	Run No.:	LB137105

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169587BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	09/08/2025