

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169575BL	WATER			Batch Number:	PB169575		Prep Date:	09/05/2025	
	Chromium	1.06	<2.5	U	5.00	P	09/08/2025	13:57	LB137113
	Copper	2.30	<5	U	10.0	P	09/08/2025	13:57	LB137113
	Nickel	1.53	<10	U	20.0	P	09/08/2025	13:57	LB137113
	Zinc	8.33	<10	U	20.0	P	09/08/2025	13:57	LB137113