

metals
- 5a -
MATRIX SPIKE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q3020
contract: VERI01 **lab code:** ACE
matrix: Water **sample id:** Q3020-01 **client id:** WATER-TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q3020-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	370		168		200	101		P
Copper	ug/L	75 - 125	148		6.46	J	150	94		P
Nickel	ug/L	75 - 125	264		10.4	J	250	101		P
Zinc	ug/L	75 - 125	633		506		100	127		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: VERINA CONSULTING GROUP, LLC **level:** low **sdg no.:** Q3020
contract: VERI01 **lab code:** ACE
matrix: Water **sample id:** Q3020-01 **client id:** WATER-TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Spiked ID:** Q3020-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Chromium	ug/L	75 - 125	366		168		200	99		P
Copper	ug/L	75 - 125	145		6.46	J	150	92		P
Nickel	ug/L	75 - 125	259		10.4	J	250	99		P
Zinc	ug/L	75 - 125	627		506		100	121		P