



Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Chromium	794	800	99	90 - 110	P	09/08/2025	12:08	LB137113
	Copper	991	1000	99	90 - 110	P	09/08/2025	12:08	LB137113
	Nickel	1930	2000	96	90 - 110	P	09/08/2025	12:08	LB137113
	Zinc	1970	2000	98	90 - 110	P	09/08/2025	12:08	LB137113



Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Chromium	10.1	10.0	101	80 - 120	P	09/08/2025	12:14	LB137113
	Copper	22.0	20.0	110	80 - 120	P	09/08/2025	12:14	LB137113
	Nickel	39.9	40.0	100	80 - 120	P	09/08/2025	12:14	LB137113
	Zinc	42.5	40.0	106	80 - 120	P	09/08/2025	12:14	LB137113

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Contract: VERI01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Chromium	1010	1000	101	90 - 110	P	09/08/2025	12:43	LB137113
	Copper	1250	1250	100	90 - 110	P	09/08/2025	12:43	LB137113
	Nickel	2480	2500	99	90 - 110	P	09/08/2025	12:43	LB137113
	Zinc	2530	2500	101	90 - 110	P	09/08/2025	12:43	LB137113
CCV02	Chromium	993	1000	99	90 - 110	P	09/08/2025	13:36	LB137113
	Copper	1250	1250	100	90 - 110	P	09/08/2025	13:36	LB137113
	Nickel	2460	2500	99	90 - 110	P	09/08/2025	13:36	LB137113
	Zinc	2500	2500	100	90 - 110	P	09/08/2025	13:36	LB137113
CCV03	Chromium	970	1000	97	90 - 110	P	09/08/2025	14:42	LB137113
	Copper	1240	1250	99	90 - 110	P	09/08/2025	14:42	LB137113
	Nickel	2430	2500	97	90 - 110	P	09/08/2025	14:42	LB137113
	Zinc	2470	2500	99	90 - 110	P	09/08/2025	14:42	LB137113
CCV04	Chromium	993	1000	99	90 - 110	P	09/08/2025	15:44	LB137113
	Copper	1260	1250	100	90 - 110	P	09/08/2025	15:44	LB137113
	Nickel	2460	2500	98	90 - 110	P	09/08/2025	15:44	LB137113
	Zinc	2500	2500	100	90 - 110	P	09/08/2025	15:44	LB137113
CCV05	Chromium	983	1000	98	90 - 110	P	09/08/2025	16:13	LB137113
	Copper	1240	1250	99	90 - 110	P	09/08/2025	16:13	LB137113
	Nickel	2430	2500	97	90 - 110	P	09/08/2025	16:13	LB137113
	Zinc	2490	2500	99	90 - 110	P	09/08/2025	16:13	LB137113