

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q3020

Contract: VERI01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	57.7	52.0	111	42	62	09/08/2025	12:27	LB137113
	Copper	-4.30	2.0	215	-18	22	09/08/2025	12:27	LB137113
	Nickel	8.14	2.0	407	-38	42	09/08/2025	12:27	LB137113
	Zinc	4.18			-40	40	09/08/2025	12:27	LB137113
ICSAB01	Chromium	586	542	108	460	624	09/08/2025	12:31	LB137113
	Copper	483	511	94	434	588	09/08/2025	12:31	LB137113
	Nickel	986	954	103	810	1100	09/08/2025	12:31	LB137113
	Zinc	1060	952	111	809	1095	09/08/2025	12:31	LB137113
ICSA	Chromium	53.2	52.0	102	42	62	09/08/2025	12:35	LB137113
	Copper	10.1	2.0	504	-18	22	09/08/2025	12:35	LB137113
	Nickel	11.9	2.0	595	-38	42	09/08/2025	12:35	LB137113
	Zinc	-4.54			-40	40	09/08/2025	12:35	LB137113
ICSAB	Chromium	673	542	124	460	624	09/08/2025	12:39	LB137113
	Copper	560	511	110	434	588	09/08/2025	12:39	LB137113
	Nickel	1050	954	110	810	1100	09/08/2025	12:39	LB137113
	Zinc	1120	952	118	809	1095	09/08/2025	12:39	LB137113