

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3024

Project: Linden Generating Station - Towers

Run No.: LB137093

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137093BS							
TDS	mg/L	100	94.0		94	1	90-110	09/05/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3024

Project: Linden Generating Station - Towers

Run No.: LB137104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137104BS							
TSS	mg/L	550	531		96	1	90-110	09/08/2025