

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

RunNo.: LB137099

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.488	0.50	98	90-110	09/05/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.494	0.5	99	90-110	09/05/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.490	0.5	98	90-110	09/05/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

RunNo.: LB137100

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.509	0.50	102	90-110	09/05/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.501	0.50	100	90-110	09/05/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.498	0.50	100	90-110	09/05/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

RunNo.: LB137105

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	09/08/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	09/08/2025