

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137083

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137083BS							
Oil and Grease	mg/L	20.0	17.0		85	1	78-114	09/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137092

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137092BS							
BOD5	mg/L	198	201		102	1	84.6-115.4	09/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137099

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137099BS							
Orthophosphate as P	mg/L	0.5	0.50		101	1	90-110	09/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137104BS							
TSS	mg/L	550	531		96	1	90-110	09/08/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137100

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169579BS							
Phosphorus, Total	mg/L	0.50	0.51		101	1	90-110	09/05/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3028

Project: Pre Treatment Plant 2025

Run No.: LB137105

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169587BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	09/08/2025