

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3029

Client ID : SSTD020680

Date Analyzed: 09/10/2025

Lab File ID: BP025704.D

Time Analyzed: 09:53

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	145300	7.755	604918	10.53	417222	14.37
UPPER LIMIT	290600	8.255	1209840	11.031	834444	14.872
LOWER LIMIT	72650	7.255	302459	10.031	208611	13.872
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	111352	7.77	431380	10.54	318751	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.