

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3029

Client ID: SSTD020680

Date Analyzed: 09/10/2025

Lab File ID: BP025704.D

Time Analyzed: 09:53

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	874394	17.172	957371	21.636	1030490	24.995
UPPER LIMIT	1748790	17.672	1914740	22.136	2060980	25.495
LOWER LIMIT	437197	16.672	478686	21.136	515245	24.495
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	699486	17.19	811066	21.63	856334	24.98

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.