



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CHEM02
Lab Code: ACE SDG No.: Q3029
Instrument ID: MSVOA_N Calibration Date/Time: 09/05/2025 11:12
Lab File ID: VN087733.D Init. Calib. Date(s): 08/21/2025 08/21/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 12:09 14:44
GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	47.14	5.718002	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.