

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN082125

Review By	John Carlone	Review On	8/22/2025 9:17:52 AM
Supervise By	Maresh Dadoda	Supervise On	8/25/2025 8:11:38 AM
SubDirectory	VN082125	HP Acquire Method	HP Processing Method 82N082125W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135214 VP135215,VP135216,VP135217,VP135218,VP135219,VP135220 VP135221

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087614.D	21 Aug 2025 09:30		JC\MD	Ok
2	VSTDCCC020	VSTDCCC020	VN087615.D	21 Aug 2025 10:04	Need ICAL	JC\MD	Not Ok
3	VN0821WBS01	VN0821WBS01	VN087616.D	21 Aug 2025 10:37		JC\MD	Not Ok
4	VN0821WBSD01	VN0821WBSD01	VN087617.D	21 Aug 2025 11:11		JC\MD	Not Ok
5	VN0821WBL01	VN0821WBL01	VN087618.D	21 Aug 2025 11:36		JC\MD	Not Ok
6	VSTDICC001	VSTDICC001	VN087619.D	21 Aug 2025 12:09	LR-10,20	JC\MD	Ok,M
7	VSTDICC005	VSTDICC005	VN087620.D	21 Aug 2025 13:08		JC\MD	Ok,M
8	VSTDICC020	VSTDICC020	VN087621.D	21 Aug 2025 13:41	method not used for #N-amyle Acetate	JC\MD	Ok,M
9	VSTDICCC050	VSTDICCC050	VN087622.D	21 Aug 2025 14:02		JC\MD	Ok,M
10	VSTDICC100	VSTDICC100	VN087623.D	21 Aug 2025 14:23		JC\MD	Ok,M
11	VSTDICC150	VSTDICC150	VN087624.D	21 Aug 2025 14:44		JC\MD	Ok,M
12	IBLK	IBLK	VN087625.D	21 Aug 2025 15:05		JC\MD	Ok
13	VSTDICV050	ICVVN082125	VN087626.D	21 Aug 2025 15:47	Icv Failed for Com.# 64 for DOD	JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN090525

Review By	Semsettin Yesilyurt	Review On	9/8/2025 9:00:38 AM
Supervise By		Supervise On	
SubDirectory	VN090525	HP Acquire Method	MSVOA_N HP Processing Method VN082125W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135381		
CCC	VP135382,VP135383		
Internal Standard/PEM	VP134956		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087732.D	05 Sep 2025 09:49		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087733.D	05 Sep 2025 11:12		JC\MD	Ok,M
3	VN0905MBL01	VN0905MBL01	VN087734.D	05 Sep 2025 11:52		JC\MD	Ok
4	VN0905WBL01	VN0905WBL01	VN087735.D	05 Sep 2025 12:13		JC\MD	Ok
5	VN0905WBS01	VN0905WBS01	VN087736.D	05 Sep 2025 12:46		JC\MD	Ok,M
6	VN0905WBSD01	VN0905WBSD01	VN087737.D	05 Sep 2025 13:07		JC\MD	Ok,M
7	Q3022-01	1413	VN087738.D	05 Sep 2025 13:39	vial A pH<2	JC\MD	Ok
8	VIBLK	VIBLK	VN087739.D	05 Sep 2025 14:00	vial A pH<2	JC\MD	Ok
9	Q3029-01	STORAGE-BLANK-SO	VN087740.D	05 Sep 2025 14:21	vial A pH<2	JC\MD	Ok
10	Q3029-02	STORAGE-BLANK-WA	VN087741.D	05 Sep 2025 14:42	vial A pH<2	JC\MD	Ok
11	Q3029-03	STORAGE-BLANK-WA	VN087742.D	05 Sep 2025 15:03	vial A pH<2	JC\MD	Ok
12	Q3029-04	STORAGE-BLANK-SAI	VN087743.D	05 Sep 2025 15:24	vial A pH<2	JC\MD	Ok
13	Q3033-01	BUR-25-0077	VN087744.D	05 Sep 2025 15:45	foamy sample no straight run.vial A pH<2	JC\MD	Ok
14	VIBLK	VIBLK	VN087745.D	05 Sep 2025 16:06		JC\MD	Ok
15	VIBLK	VIBLK	VN087746.D	05 Sep 2025 16:48		JC\MD	Ok
16	VSTDCCC050	VSTDCCC050EC	VN087747.D	05 Sep 2025 17:09		JC\MD	Ok,M

M : Manual Integration