

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3029
 Lab File ID: VN087733.D Date Analyzed: 09/05/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	206857	13.77				
UPPER LIMIT	413714	14.27				
LOWER LIMIT	103429	13.27				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	162177	13.77				
STORAGE-BLANK-WATER-REF-5	153634	13.77				
STORAGE-BLANK-WATER-REF-4	152855	13.77				
STORAGE-BLANK-SAMPLE-MANAGEMENT	148118	13.77				
VN0905WBL01	156701	13.77				
VN0905WBS01	194094	13.77				
VN0905WBSD01	200325	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.