


LB137126

INSTRUMENT ID: CV1

Reviewed By: jaswal
On: 9/10/2025 11:43:30 PM
Inst Id: CV1
LB: LB137126

7470A



A= 0.0000e+000

B= 3.0617e-004

C= -4.9050e-002

Rho= 0.9991191

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.028	-0.028	69	0.000	69	0				
0.2	0.200	0.218	0.018	872	0.0 %	872	0				
2.5	2.500	2.455	-0.045	8178	0.0 %	8178	0				
5.0	5.000	5.252	0.252	17313	0.0 %	17313					
7.5	7.500	7.231	-0.269	23779	0.0 %	23779					
10.0	10.000	10.072	0.072	33058	0.0 %	33058					

LB137126 INSTRUMENT ID CV:1

Method: 7470A Operator: Admin Date of Analysis: 10 Sep 2025 09:19:57

Type	Sample ID	Extended ID	Conc.	µ Abs	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	69	PPB	Std	0.0000	7470A	10 Sep 2025 09:23:11
S	0.2 - 1	50.2	-	872	PPB	Std	0.2000	7470A	10 Sep 2025 09:28:08
S	2.5 - 1	52.5	-	8178	PPB	Std	2.5000	7470A	10 Sep 2025 09:30:25
S	5.0 - 1	52.5	-	17313	PPB	Std	5.0000	7470A	10 Sep 2025 09:32:41
S	7.5 - 1	55.0	-	23779	PPB	Std	7.5000	7470A	10 Sep 2025 09:35:00
S	10.0 - 1	57.5	-	33058	PPB	Std	10.0000	7470A	10 Sep 2025 09:40:29
U	ICV24 - 1	ICV24	4.3117	14243	PPB	SMPL	-	7470A	10 Sep 2025 09:48:25
U	ICB24 - 1	ICB24	0.0048	176	PPB	SMPL	-	7470A	10 Sep 2025 09:56:19
U	CCV70 - 1	CCV70	5.1279	16909	PPB	SMPL	-	7470A	10 Sep 2025 10:03:56
U	CCB70 - 1	CCB70	-0.0224	87	PPB	SMPL	-	7470A	10 Sep 2025 10:09:02
U	CRA - 1	CRA	0.2139	859	PPB	SMPL	-	7470A	10 Sep 2025 10:24:16
U	HighStd - 1	HighStd	9.7961	32156	PPB	SMPL	-	7470A	10 Sep 2025 10:33:59
U	ChkStd - 1	ChkStd	8.2888	27233	PPB	SMPL	-	7470A	10 Sep 2025 10:36:16
U	PB169614BL - 1	PBW	-0.0298	63	PPB	SMPL	-	7470A	10 Sep 2025 10:41:14
U	PB169614BS - 1	LCSW	4.4801	14793	PPB	SMPL	-	7470A	10 Sep 2025 10:46:25
U	Q3030-01 - 1	BLEECKER-082825HU01-TCLP	-0.1443	-311	PPB	SMPL	-	7470A	10 Sep 2025 10:48:41
U	Q3032-02 - 1	SOIL-PILE-1	-0.1133	-210	PPB	SMPL	-	7470A	10 Sep 2025 10:51:00
U	Q3032-05 - 1	SOIL-PILE-2	-0.1069	-189	PPB	SMPL	-	7470A	10 Sep 2025 10:53:17
U	Q3050-01 - 1	72-11939	-0.0791	-98	PPB	SMPL	-	7470A	10 Sep 2025 10:55:33
U	Q3051-02 - 1	VNJ-223	-0.0659	-55	PPB	SMPL	-	7470A	10 Sep 2025 10:57:49
U	CCV71 - 1	CCV71	4.7951	15822	PPB	SMPL	-	7470A	10 Sep 2025 11:05:11
U	CCB71 - 1	CCB71	-0.0080	134	PPB	SMPL	-	7470A	10 Sep 2025 11:07:26
U	Q3051-02DUP - 1	VNJ-223DUP	-0.0126	119	PPB	SMPL	-	7470A	10 Sep 2025 11:09:44
U	Q3051-02MS - 1	VNJ-223MS	4.4991	14855	PPB	SMPL	-	7470A	10 Sep 2025 11:14:39
U	Q3051-02MSD - 1	VNJ-223MSD	3.9856	13178	PPB	SMPL	-	7470A	10 Sep 2025 11:20:02
U	PB169616BL - 1	PBW	-0.0460	10	PPB	SMPL	-	7470A	10 Sep 2025 11:24:57
U	PB169616BS - 1	LCSW	3.3855	11218	PPB	SMPL	-	7470A	10 Sep 2025 11:30:01
U	Q3033-01 - 1	BUR-25-0077	0.4803	1729	PPB	SMPL	-	7470A	10 Sep 2025 11:32:19
U	Q3033-01DUP - 1	BUR-25-0077DUP	0.2152	863	PPB	SMPL	-	7470A	10 Sep 2025 11:43:03
U	Q3033-01MS - 1	BUR-25-0077MS	4.3934	14510	PPB	SMPL	-	7470A	10 Sep 2025 11:45:20
U	Q3033-01MSD - 1	BUR-25-0077MSD	3.4324	11371	PPB	SMPL	-	7470A	10 Sep 2025 11:47:37
U	Q3036-03DLX25 - 1	EGR-2	-0.1614	-367	PPB	SMPL	-	7470A	10 Sep 2025 11:49:58
U	CCV72 - 1	CCV72	4.6411	15319	PPB	SMPL	-	7470A	10 Sep 2025 12:14:07
U	CCB72 - 1	CCB72	-0.0490	0	PPB	SMPL	-	7470A	10 Sep 2025 12:31:05
U	Q3045-02 - 1	DNAPL	-0.2845	-769	PPB	SMPL	-	7470A	10 Sep 2025 12:58:24
U	P169589TB - 1	P169589TB	-0.0490	0	PPB	SMPL	-	7470A	10 Sep 2025 13:03:47
U	Q3051-02LX5 - 1		-0.2517	-662	PPB	SMPL	-	7470A	10 Sep 2025 13:06:40
U	Q3051-02A - 1		3.9134	12942	PPB	SMPL	-	7470A	10 Sep 2025 13:09:30
U	Q3033-01LX5 - 1		-0.1574	-354	PPB	SMPL	-	7470A	10 Sep 2025 13:12:09
U	Q3033-01A - 1		3.6740	12160	PPB	SMPL	-	7470A	10 Sep 2025 13:14:37
U	CCV73 - 1	CCV73	4.7256	15595	PPB	SMPL	-	7470A	10 Sep 2025 13:20:48
U	CCB73 - 1	CCB73	0.0107	195	PPB	SMPL	-	7470A	10 Sep 2025 13:26:03