

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CHA Companies, Inc.

SDG No.: Q3030

Contract: CLOU03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169614BS Mercury	ug/L	4.0	4.48		112	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CHA Companies, Inc.

SDG No.: Q3030

Contract: CLOU03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169617BS							
Arsenic	ug/L	4000	3810		95	80 - 120	P
Barium	ug/L	1000	887		89	80 - 120	P
Cadmium	ug/L	1000	932		93	80 - 120	P
Chromium	ug/L	2000	1950		98	80 - 120	P
Lead	ug/L	5000	4690		94	80 - 120	P
Selenium	ug/L	10000	9720		97	80 - 120	P
Silver	ug/L	380	361		95	80 - 120	P