

Initial and Continuing Calibration Verification

Client: Zuccaro Inc.

SDG No.: Q3032

Project: 3 Coal Ave., Morristown, NJ

RunNo.: LB137132

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	90-110	09/10/2025
Sample ID: Cyanide	CCV1	mg/L	0.25	0.25	100	90-110	09/10/2025
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	09/10/2025
Sample ID: Cyanide	CCV3	mg/L	0.25	0.25	100	90-110	09/10/2025
Sample ID: Cyanide	CCV4	mg/L	0.26	0.25	104	90-110	09/10/2025