



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	105	100	105	65 - 135	P	09/08/2025	12:22	LB137113
	Antimony	47.9	50.0	96	65 - 135	P	09/08/2025	12:22	LB137113
	Arsenic	21.0	20.0	105	65 - 135	P	09/08/2025	12:22	LB137113
	Barium	100	100	100	65 - 135	P	09/08/2025	12:22	LB137113
	Beryllium	6.37	6.0	106	65 - 135	P	09/08/2025	12:22	LB137113
	Cadmium	5.92	6.0	99	65 - 135	P	09/08/2025	12:22	LB137113
	Calcium	2090	2000	104	65 - 135	P	09/08/2025	12:22	LB137113
	Chromium	10.3	10.0	103	65 - 135	P	09/08/2025	12:22	LB137113
	Cobalt	29.8	30.0	99	65 - 135	P	09/08/2025	12:22	LB137113
	Copper	22.0	20.0	110	65 - 135	P	09/08/2025	12:22	LB137113
	Iron	106	100	106	65 - 135	P	09/08/2025	12:22	LB137113
	Lead	11.7	12.0	98	65 - 135	P	09/08/2025	12:22	LB137113
	Magnesium	2170	2000	108	65 - 135	P	09/08/2025	12:22	LB137113
	Manganese	22.5	20.0	113	65 - 135	P	09/08/2025	12:22	LB137113
	Nickel	39.6	40.0	99	65 - 135	P	09/08/2025	12:22	LB137113
	Potassium	1930	2000	97	65 - 135	P	09/08/2025	12:22	LB137113
	Selenium	19.7	20.0	99	65 - 135	P	09/08/2025	12:22	LB137113
	Silver	10.1	10.0	101	65 - 135	P	09/08/2025	12:22	LB137113
	Sodium	1980	2000	99	65 - 135	P	09/08/2025	12:22	LB137113
	Thallium	38.7	40.0	97	65 - 135	P	09/08/2025	12:22	LB137113
	Vanadium	41.8	40.0	104	65 - 135	P	09/08/2025	12:22	LB137113
	Zinc	42.4	40.0	106	65 - 135	P	09/08/2025	12:22	LB137113
CRA	Mercury	0.16	0.2	81	70 - 130	CV	09/08/2025	14:46	LB137106