

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169595BS							
Aluminum	mg/Kg	100	95.9		96	80 - 120	P
Antimony	mg/Kg	40.0	38.6		96	80 - 120	P
Arsenic	mg/Kg	40.0	38.1		95	80 - 120	P
Barium	mg/Kg	10.0	9.15		92	80 - 120	P
Beryllium	mg/Kg	10.0	9.51		95	80 - 120	P
Cadmium	mg/Kg	10.0	9.33		93	80 - 120	P
Calcium	mg/Kg	50.0	48.3	J	97	80 - 120	P
Chromium	mg/Kg	20.0	19.3		96	80 - 120	P
Cobalt	mg/Kg	10.0	9.44		94	80 - 120	P
Copper	mg/Kg	15.0	14.9		99	80 - 120	P
Iron	mg/Kg	150	143		95	80 - 120	P
Lead	mg/Kg	50.0	46.0		92	80 - 120	P
Magnesium	mg/Kg	100	94.3	J	94	80 - 120	P
Manganese	mg/Kg	10.0	9.46		95	80 - 120	P
Nickel	mg/Kg	25.0	23.5		94	80 - 120	P
Potassium	mg/Kg	500	487		97	80 - 120	P
Selenium	mg/Kg	100	97.7		98	80 - 120	P
Silver	mg/Kg	3.8	3.58		94	80 - 120	P
Sodium	mg/Kg	150	154		103	80 - 120	P
Thallium	mg/Kg	100	94.4		94	80 - 120	P
Vanadium	mg/Kg	15.0	14.3		95	80 - 120	P
Zinc	mg/Kg	10.0	9.91		99	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Zuccaro Inc.

SDG No.: Q3032

Contract: ZUCC01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169605BS Mercury	mg/Kg	0.25	0.23		92	80 - 120	CV